

# Realistic Head Phantom for Evaluation of Brain Stroke Localization Methods Using 3D Printer

Juneseok Lee · Jihoon Bang · Jaehoon Choi\*

## Abstract

In this paper, a brain phantom for evaluating brain stroke localization is proposed. To evaluate brain stroke localization, a phantom imitating three-dimensional (3D) simulation environment is needed. Mold for the proposed phantom was printed by a 3D printer and the interior of the phantom consists of 5 different brain tissue materials. Each of the brain tissue materials has the conductivity and permittivity similar to those of the International Commission on Non-Ionizing Radiation Protection (ICNIRP) standards for a frequency band from 0.5 to 2 GHz.

**Key Words:** Brain, Phantom, Stroke Localization, 3D Printer.

## I. INTRODUCTION

The number of deaths caused by a brain stroke has rapidly increased over the years. The interest in microwave brain stroke localization has drawn people's interest because of its low cost and portability [1]. Several research groups have reported microwave imaging systems or devices [2–6]. To validate the effectiveness of microwave imaging methods, a phantom imitating human head in electrical and anatomical aspects is required [6–8]. General human phantoms for specific absorption rate (SAR) tests aim to mimic human body in a narrow frequency band which is not suitable for a microwave imaging method requiring the broadband (0.5–2 GHz) characteristics [9, 10]. Several wideband phantoms have a simple rectangular shape or are too homogeneous, which makes them too simple to model an actual brain stroke [5, 8]. There is a head phantom using a commercial plastic skull model with layered tissue but the model is too simple to match well with actual brain for electromagnetic (EM)

analysis [8]. In this paper, a head phantom for evaluation of a brain stroke localization is proposed. Firstly, molds for the proposed phantom was printed by 3D printer to imitate EM simulation setup in structural aspect. After that, 4 different brain tissue materials (spinal cord, cerebellum, white matter, and cerebrospinal fluid) were fabricated with polyethylene (PE) powder, agar, Tx-151, NaCl, NaN<sub>3</sub>, and distilled water as infilled materials of the phantom. Blood material was additionally fabricated for simulating stroke event with the same ingredients aforementioned. The electrical properties of the brain tissue materials and blood material in the frequency band from 0.5 to 2 GHz are very close to those suggested in [11]. The brain phantom was fabricated with the molds and 4 brains tissue materials. At the last stage, blood material is implanted as it is demanded.

## II. FABRICATION

### 1. Molds for the Brain Phantom

Manuscript received August 12, 2016 ; Revised October 12, 2016 ; Accepted October 17, 2016. (ID No. 20160812-026J)

Department of Electronics and Computer Engineering, Hanyang University, Seoul, Korea

\*Corresponding Author: Jaehoon Choi (e-mail: [choijh@hanyang.ac.kr](mailto:choijh@hanyang.ac.kr))

This is an Open-Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<http://creativecommons.org/licenses/by-nc/3.0>) which permits unrestricted non-commercial use, distribution, and reproduction in any medium, provided the original work is properly cited.

© Copyright The Korean Institute of Electromagnetic Engineering and Science. All Rights Reserved.

Dimensions for the molds of a phantom (height, 265 mm; sagittal width, 260 mm; coronal width, 197 mm; and thickness of mold, 8–12 mm) are the same as brain CAD data provided by ANSYS HFSS (ver. 14). However, the CAD data could not be applied directly to the phantom molds since the data had holes and overlapping parts. The ANSYS CAD data was modified to enclose each brain tissue and to eliminate overlapped parts. In addition, the oval and nasal cavities were eliminated. The brain model was printed as molds by MakerBot 3D printer whose resolution was 0.1 mm [12]. Fig. 1(a) shows the 3D printer. Even though the real brain consists of a lot of tissues,

the brain was subdivided into 4 major tissues (spinal cord, cerebellum, white matter, and cerebrospinal fluid) in our head phantom model. The printed molds consist of outer parts and inner molds, including white matter, cerebellum, and spinal cord, as shown in Fig. 1(b) and (c). The outer molds were used as the outer shell of the brain phantom, while the inner molds were used to separate and shape the brain tissue materials, and were extracted after the tissues solidified. The inner molds were used only for shaping and locating the brain tissue materials. Spinal cord, cerebellum, and white matter among the subdivided 4 major brain tissues were embedded with supporters to main the proper locations. The filament used for making the molds is polylactic acid (PLA) and its relative permittivity is about 2.0.

## 2. Recipes for the Blood and Brain Tissue Materials

Recipes for the 4 brain tissue materials and blood material were developed such that the electrical properties of the phantom should be similar to those in [11]. The fabricated blood material was used for simulating a blood clot caused by a brain stroke. Each brain tissue material consists of PE powder, agar, TX-151, NaCl, NaN<sub>3</sub>, and distilled water. Distilled water ( $\epsilon_r \approx 80$ ) is basic solution. PE powder and NaCl control permittivity and conductivity, respectively. Fine PE particles produce dispersive dielectric constants along the frequency band and ionized NaCl adjusts conductivity. Precipitation of PE powder and NaCl result in changing electrical properties on the materials. To prevent the precipitation for long time, solidification is needed. However, the precipitation is happened during the transient period from the liquid state to the solid state. Gelification can prevent the precipitation during the transient period.


TX-151 and agar were used for gelification and solidification, respectively. NaN<sub>3</sub> was added as a preservative. Ingredient proportion of each brain tissue material is shown in Table 1. Fig. 2 shows the fabrication procedure of brain tissue materials and the detailed procedure is described below.

1) Add NaCl and NaN<sub>3</sub> to distilled water in a mixing ball.


Before adding agar and TX-151, heat the mixture to 40°C

Table 1. Ingredient proportion of each brain tissue material (unit, g)


	H <sub>2</sub> O	Agar	Tx-151	PE	NaCl	NaN <sub>3</sub>
Blood	1,000	80	35	75	34	1
Cerebrospinal fluid	1,000	80	35	100	48	1
Cerebellum	1,000	80	35	125	29	1
Spinal cord	1,000	80	35	550	7	1
White matter	1,000	80	35	450	8	1



(a)



(b)



(c)

Fig. 1. The 3D printed molds for the phantom: (a) 3D printer, (b) 3D CAD data of the molds, and (c) photo of the molds.





Fig. 2. Fabrication procedure of a brain tissue material.


to dissolve the agar and TX-151.

- 2) The mixture was heated up to 70°C. A stirrer should be continuously operating while heating.
- 3) Once temperature of the mixture reaches 70°C, PE powder was added and turn off the heater.
- 4) Keep stirring the mixture until all PE powder was evenly mixed.
- 5) The brain tissue material was poured into the mold and cooled down for 24 hours.

Permittivity and conductivity of the fabricated brain tissue materials and blood material were measured after 24 hours with Agilent 85070E dielectric probe kit and 8719ES network analyzer. The measured permittivity and conductivity values of the brain tissue materials consisting brain phantom should have less than 10% error comparing to those suggested in [11]. The measured data and standard data were compared in Fig. 3. Relative permittivities of the brain tissues are varied less than 10% in the frequency band from 0.5 to 2 GHz the dispersive characteristic of the conductivity of fabricated brain tissue materials is very similar to that of real brain tissue given in [11]. The maximum error in the conductivity was observed in blood material



(a)



(b)

Fig. 3. Conductivities and permittivities of the phantom: (a) relative permittivities and (b) conductivities.

at 2 GHz and is less than 10%.

### 3. Brain Phantom with Stroke

To evaluate imaging method for brain stroke localization, a 4-layered brain phantom with blood material simulating blood clot was needed.

To fabricate the 4-layered phantom, cerebrospinal fluid brain tissue materials was poured into the outer shell while other inner molds were remaining. After the material was firmly solidified, the inner molds were taken off as shown in Fig. 4(a). To fabricate other layers, the aforementioned procedure was repeatedly performed in the following order: white matter, cerebellum, and spinal cord as shown in Fig. 4(b) and (c). The halves of the phantom were put together as shown in Fig. 4(d). As the last procedure, blood tissue material was shaped and located at the position where the blood clot, simulating brain stroke, was placed in the EM simulation as shown in Figs. 5 and 6. Fig. 6 shows cross section of the final manufactured brain phantom with blood material.

## III. CONCLUSION






Fig. 4. Process of fabricating the phantom: (a) exterior mold with cerebrospinal fluid material, (b) after white matter material was filled and cooling with cerebellum mold and spinal cord mold, (c) cerebellum mold and spinal cord mold extracted, and (d) the final manufactured brain phantom.




Fig. 5. The modified 3D CAD data with stroke event.

In this paper a brain phantom for evaluating brain stroke localization is proposed. Outer shell of the proposed phantom was




Fig. 6. Cross section of the phantom: (a) upper area and (b) lower area.

printed by 3D printer and 4 different brain tissue and blood materials were fabricated and located in the printed molds. The brain phantom imitated a brain model in EM simulation in terms of electrical properties similar to those provide in [11] and structural shape. Thus the proposed phantom can be used to validate the effectiveness of microwave imaging methods and provide a means of comparison between measurement and simulation for evaluating brain stroke localization.

This research was supported by the Ministry of Science, ICT and Future Planning (MSIP), Korea, under the Information Technology Research Center support program (No. IITP-2016-H8601-16-1005) supervised by the Institute for Information & Communications Technology Promotion (IITP).

## REFERENCES

- [1] The Internet Stroke Center [Online]. Available: <http://www.strokecenter.org/>.
- [2] X. Li and S. C. Hagness, "A confocal microwave imaging algorithm for breast cancer detection," *IEEE Microwave and Wireless Components Letters*, vol. 11, no. 3, pp. 130–132, 2001.
- [3] J. C. Bolomey and C. Pichot, "Microwave tomography: from theory to practical imaging systems," *International Journal of Imaging Systems and Technology*, vol. 2, no. 2, pp. 144–156, 1990.
- [4] S. Y. Semenov and D. R. Corfield, "Microwave tomography for brain imaging: feasibility assessment for stroke detection," *International Journal of Antennas and Propagation*, vol. 2008, article ID. 254830, 2008.
- [5] L. Jofre, M. S. Hawley, A. Broquetas, E. de Los Reyes, M. Ferrando, and A. R. Elias-Fuste, "Medical imaging with a microwave tomographic scanner," *IEEE Transactions on Biomedical Engineering*, vol. 37, no. 3, pp. 303–312, 1990.
- [6] S. Mustafa, B. Mohammed, and A. Abbosh, "Novel pre-processing techniques for accurate microwave imaging of human brain," *IEEE Antennas and Wireless Propagation Letters*, vol. 12, pp. 460–463, 2013.
- [7] B. J. Mohammed, A. Abbosh, S. Mustafa, and D. Ireland, "Microwave system for head imaging," *IEEE Transactions on Instrumentation and Measurement*, vol. 63, no. 1, pp. 117–123, 2014.
- [8] A. Abbosh, B. Henin, and P. Sharpe, "Head phantom for testing microwave systems for head imaging," in *Proceedings of Cairo International Biomedical Engineering Conference (CIBEC)*, Cairo, 2012, pp. 191–193.
- [9] S. Lee, W. Seo, K. Kwon, and J. Choi, "The study on implementation of a semi-solid flat phantom with equivalent electrical properties to whole human body at MICS and ISM band," *Journal of Korean Institute of Electromagnetic Engineering and Science*, vol. 23, no. 1, pp. 101–107, 2012.
- [10] P. J. Dimbylow and S. M. Mann, "SAR calculations in an anatomically realistic model of the head for mobile communication transceivers at 900 MHz and 1.8 GHz," *Physics in Medicine and Biology*, vol. 39, no. 10, pp. 1537–1553, 1994.
- [11] C. Gabriel, "Compilation of the dielectric properties of body tissues at RF and microwave frequencies," Occupational and Environmental Health Directorate, Armstrong Laboratory, Brooks Air Force Base, TX, Report No. AL/OE-TR-1996-0037, 1996.
- [12] MakerBot Replicator 5th Generation 3D Printer [Online]. Available: [http://download.makerbot.com/replicator/MB\\_Replicator\\_UserManual.pdf](http://download.makerbot.com/replicator/MB_Replicator_UserManual.pdf).

#### Juneseok Lee



received a B.S. in Electronics and Computer Engineering from Chungnam National University, Daejeon, Korea, in 2010. He also received an M.S. from the Department of Electrical Engineering at The University of Queensland, Brisbane, Australia, in 2011. He is currently undertaking his Ph.D. research at Hanyang University, Seoul, Korea. He is interested in RF components including antenna design, wireless communication systems, wireless power transferring, and wireless body area networks. Currently, his research is focused on microwave imaging system for brain stroke localization.

#### Jaehoon Choi



received a B.S. from Hanyang University, Korea, and M.S. and Ph.D. from Ohio State University, Ohio, in 1980, 1986, and 1989, respectively. From 1989–1991, he was a research analyst with the Telecommunication Research Center at Arizona State University, Tempe, Arizona. He worked for Korea Telecom as a team leader of the Satellite Communication Division from 1991 to 1995. Since 1995, he has been a professor in the Department of Electronic Engineering at Hanyang University, Korea. He has published more than 200 peer-reviewed journal articles and has contributed to numerous conference proceedings. He also holds over 50 patents. His research interests include antennas, microwave circuit design, and EMC. Currently, his research is mainly focused on the design of compact, multi-band antennas for mobile wireless communication, and other biomedical applications.

#### Jihoon Bang



received a B.S. in Information and Communication Engineering from Dongyang Mirae University, Seoul, Korea, in 2016. He is currently working toward the unified course of the M.S. and the Ph.D. degree in electronics and computer engineering at the Hanyang University, Seoul, Korea. He is interested in antenna design and microwave circuit design. Currently, his research is focused on microwave imaging antenna for brain stroke localization.